



TRUSTI®

Digital Refractometer Readings:

To test Brix measure in colostrum (S01):

- Turn ON (READ button)
- Choose the SCALE S01
- Drop the colostrum sample into the sample tank and close the lid
- Press READ to activate the measuring
- A reading $>22\%$ correlates to a high level of antibodies/immunoglobulins in the colostrum and is considered good quality for first feeding

